

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042924\
 Data File : VU058816.D
 Acq On : 29 Apr 2024 14:13
 Operator : MD/SY
 Sample : P2251-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR3

Quant Time: Apr 30 06:12:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 30 06:10:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	65674	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	62428	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	29043	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18386	3.559	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.200%
7) Chloroethane-d5	1.904	69	15798	4.058	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.200%
11) 1,1-Dichloroethene-d2	2.557	65	11289	4.341	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	4.608	46	41134	45.322	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.640%
24) Chloroform-d	5.049	84	53154	5.033	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.692	65	28208	5.511	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
32) Benzene-d6	5.718	84	94596	4.761	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
36) 1,2-Dichloropropane-d6	6.679	67	26784	4.519	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.400%
41) Toluene-d8	7.891	98	82704	4.650	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	8.174	79	10879	5.042	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.800%
46) 2-Hexanone-d5	8.624	63	30402	47.547	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.100%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	17345	4.805	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	28308	5.592	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042924\
Data File : VU058816.D
Acq On : 29 Apr 2024 14:13
Operator : MD/SY
Sample : P2251-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR3

Quant Time: Apr 30 06:12:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 30 06:10:03 2024
Response via : Initial Calibration

