

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\  
Data File : VU023534.D  
Acq On : 30 Apr 2018 18:18  
Operator : MD/SY  
Sample : VU0430WBL01  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0430WBL01

Quant Time: May 01 05:27:42 2018  
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Tue May 01 05:06:52 2018  
Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|--------------------|------|------|----------|------|-------|----------|
| 1) Fluorobenzene   | 5.74 | 96   | 50168    | 1.00 | ug/l  | 0.00     |

## System Monitoring Compounds

|                            |       |     |          |      |        |      |
|----------------------------|-------|-----|----------|------|--------|------|
| 57) 4-Bromofluorobenzene   | 10.31 | 95  | 13392    | 0.82 | ug/l   | 0.00 |
| Spiked Amount              | 1.000 |     | Recovery | =    | 82.00% |      |
| 68) 1,2-Dichlorobenzene-d4 | 11.87 | 152 | 13174    | 0.83 | ug/l   | 0.00 |
| Spiked Amount              | 1.000 |     | Recovery | =    | 83.00% |      |

| Target Compounds | Ovalue |
|------------------|--------|
|------------------|--------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\  
Data File : VU023534.D  
Acq On : 30 Apr 2018 18:18  
Operator : MD/SY  
Sample : VU0430WBL01  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0430WBL01

Quant Time: May 01 05:27:42 2018  
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Tue May 01 05:06:52 2018  
Response via : Initial Calibration

