

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\
 Data File : VU023536.D
 Acq On : 30 Apr 2018 19:30
 Operator : MD/SY
 Sample : J2241-04DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1213DL

Quant Time: May 01 05:33:29 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue May 01 05:06:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	53479	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	13647	0.78	ug/l	0.00
Spiked Amount 1.000			Recovery	=	78.00%	
68) 1,2-Dichlorobenzene-d4	11.87	152	14203	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	
Target Compounds						Ovalue
27) Chloroform	4.68	83	92034	3.66	ug/l	100
44) Bromodichloromethane	6.76	83	51005	2.86	ug/l	99
55) Dibromochloromethane	8.48	129	118759	10.75	ug/l	98
67) Bromoform	9.96	173	58075	10.06	ug/l	99
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\
Data File : VU023536.D
Acq On : 30 Apr 2018 19:30
Operator : MD/SY
Sample : J2241-04DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1213DL

Quant Time: May 01 05:33:29 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 01 05:06:52 2018
Response via : Initial Calibration

