

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\
 Data File : VU023537.D
 Acq On : 30 Apr 2018 19:59
 Operator : MD/SY
 Sample : J2241-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1220

Quant Time: May 01 05:35:14 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue May 01 05:06:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	46908	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	12372	0.81	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.00%	
68) 1,2-Dichlorobenzene-d4	11.87	152	12494	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	
Target Compounds						
						Ovalue
2) Dichlorodifluoromethane	1.21	85	754018	49.51	ug/l	100
3) Chloromethane	1.33	50	449415	25.47	ug/l	100
4) Vinyl Chloride	1.40	62	798539	47.17	ug/l	98
5) Bromomethane	1.63	94	363389	40.15	ug/l	100
6) Chloroethane	1.70	64	412579	43.78	ug/l	99
7) Trichlorofluoromethane	1.89	101	799117	41.12	ug/l	99
9) 1,1-Dichloroethene	2.29	96	229668	20.72	ug/l	99
10) Iodomethane	2.42	142	15314	1.16	ug/l	95
15) Methylene Chloride	2.70	84	193067	15.29	ug/l	99
16) trans-1,2-Dichloroethene	2.99	96	187314	15.77	ug/l	100
17) 1,1-Dichloroethane	3.45	63	453258	19.38	ug/l	100
21) 2,2-Dichloropropane	4.23	77	79559	4.29	ug/l	91
22) cis-1,2-Dichloroethene	4.23	96	127946	9.96	ug/l	95
26) Bromochloromethane	4.55	128	20787	4.09	ug/l	99
28) 1,1,1-Trichloroethane	4.91	97	223261	12.35	ug/l	99
29) 1,1-Dichloropropene	5.13	75	280551	15.82	ug/l	99
30) Carbon Tetrachloride	5.13	117	261919	17.23	ug/l	99
36) 1,2-Dichloroethane	5.41	62	46580	3.13	ug/l	99
37) Trichloroethene	6.19	130	99925	7.06	ug/l	99
38) 1,2-Dichloropropane	6.44	63	52677	3.64	ug/l	98
43) Dibromomethane	6.56	93	41858	6.73	ug/l	99
52) 1,1,2-Trichloroethane	8.07	97	168792	19.08	ug/l	98
53) 1,3-Dichloropropane	8.25	76	68743	4.37	ug/l	98
58) Tetrachloroethene	8.23	164	119575	9.27	ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.21	131	200402	18.11	ug/l	100
71) 1,2,3-Trichloropropane	10.50	75	114447	16.04	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_U
ClientSampleId :
1220

Quant Time: May 01 05:35:14 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 01 05:06:52 2018
Response via : Initial Calibration

