

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\
 Data File : VU023538.D
 Acq On : 30 Apr 2018 20:28
 Operator : MD/SY
 Sample : J2241-05DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1220DL

Quant Time: May 01 05:39:53 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue May 01 05:06:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	44290	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	11134	0.77	ug/l	0.00
Spiked Amount 1.000			Recovery	=	77.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	11769	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	

Target Compounds

						Ovalue
2) Dichlorodifluoromethane	1.20	85	181089	12.59	ug/l	98
3) Chloromethane	1.33	50	116306	6.98	ug/l	100
4) Vinyl Chloride	1.40	62	190470	11.92	ug/l	99
5) Bromomethane	1.63	94	91711	10.73	ug/l	99
6) Chloroethane	1.70	64	100493	11.29	ug/l	99
7) Trichlorofluoromethane	1.89	101	192958	10.52	ug/l	100
9) 1,1-Dichloroethene	2.29	96	55631	5.32	ug/l	99
15) Methylene Chloride	2.71	84	48710	4.09	ug/l	97
16) trans-1,2-Dichloroethene	2.99	96	46803	4.17	ug/l	97
17) 1,1-Dichloroethane	3.45	63	112368	5.09	ug/l	100
21) 2,2-Dichloropropane	4.23	77	19360	1.11	ug/l	94
22) cis-1,2-Dichloroethene	4.23	96	32266	2.66	ug/l	93
26) Bromochloromethane	4.55	128	4946	1.03	ug/l	97
28) 1,1,1-Trichloroethane	4.91	97	53322	3.13	ug/l	99
29) 1,1-Dichloropropene	5.13	75	67873	4.05	ug/l	99
30) Carbon Tetrachloride	5.13	117	60919	4.24	ug/l	100
36) 1,2-Dichloroethane	5.41	62	11443	0.81	ug/l	98
37) Trichloroethene	6.18	130	24488	1.83	ug/l	97
38) 1,2-Dichloropropane	6.44	63	13037	0.96	ug/l	94
43) Dibromomethane	6.56	93	10206	1.74	ug/l	98
52) 1,1,2-Trichloroethane	8.07	97	41136	4.93	ug/l	99
53) 1,3-Dichloropropane	8.25	76	16318	1.10	ug/l	98
58) Tetrachloroethene	8.23	164	28422	2.33	ug/l	96
60) 1,1,1,2-Tetrachloroethane	9.21	131	47481	4.54	ug/l	99
71) 1,2,3-Trichloropropane	10.50	75	28401	4.22	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU043018\
Data File : VU023538.D
Acq On : 30 Apr 2018 20:28
Operator : MD/SY
Sample : J2241-05DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1220DL

Quant Time: May 01 05:39:53 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\524U043018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 01 05:06:52 2018
Response via : Initial Calibration

