

Data File : VU031656.D
Acq On : 01 May 2019 00:39
Operator : JC/SP
Sample : K2604-05
Misc : 5.08g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: May 01 05:01:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 01 04:42:38 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	846257	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	804652	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	323368	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	202999	28.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	57.32%#
7) Chloroethane-d5	1.68	69	123820	22.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	45.24%#
11) 1,1-Dichloroethene-d2	2.24	63	305395	21.65	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	43.30%#
21) 2-Butanone-d5	4.21	46	343417	53.52	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	53.52%
24) Chloroform-d	4.64	84	349497	26.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	53.88%#
26) 1,2-Dichloroethane-d4	5.31	65	254705	30.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	60.12%#
32) Benzene-d6	5.33	84	682391	27.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	55.24%#
36) 1,2-Dichloropropane-d6	6.33	67	253401	29.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	58.30%#
41) Toluene-d8	7.56	98	553774	25.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	51.46%#
43) trans-1,3-Dichloropropene-	7.85	79	94846	25.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.84%#
47) 2-Hexanone-d5	8.34	63	220610	58.63	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	58.63%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	307157	26.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	52.66%#
64) 1,2-Dichlorobenzene-d4	11.86	152	180191	26.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	53.00%#

Target Compounds					Qvalue
16) Methylene chloride	2.69	84	5904m	0.758	ug/L
34) Trichloroethene	6.19	95	8038	1.121	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU031656.D
Acq On : 01 May 2019 00:39
Operator : JC/SP
Sample : K2604-05
Misc : 5.08g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: May 01 05:01:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 01 04:42:38 2019
Response via : Initial Calibration

