

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU043019\
 Data File : VU031633.D
 Acq On : 30 Apr 2019 14:55
 Operator : JC/SP
 Sample : K2606-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: May 01 11:14:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 01 01:38:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	619326	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	598522	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	206567	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	306269	59.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	118.18%
7) Chloroethane-d5	1.68	69	224982	56.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.34%
11) 1,1-Dichloroethene-d2	2.27	63	454900	44.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.14%
21) 2-Butanone-d5	4.19	46	480153	102.24	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.24%
24) Chloroform-d	4.64	84	494871	52.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.26%
26) 1,2-Dichloroethane-d4	5.31	65	348708	56.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.46%
32) Benzene-d6	5.34	84	971578	52.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.72%
36) 1,2-Dichloropropane-d6	6.32	67	341704	52.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.70%
41) Toluene-d8	7.56	98	786751	49.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.30%
43) trans-1,3-Dichloropropene-	7.85	79	132266	47.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.30%
47) 2-Hexanone-d5	8.31	63	264139	94.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.38%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	422626	48.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	223902	51.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU043019\
Data File : VU031633.D
Acq On : 30 Apr 2019 14:55
Operator : JC/SP
Sample : K2606-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: May 01 11:14:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 01 01:38:49 2019
Response via : Initial Calibration

