

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
 Data File : VU043423.D
 Acq On : 29 Apr 2021 14:27
 Operator : SY/MD
 Sample : M2197-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YARS9

Quant Time: Apr 30 07:28:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 30 07:27:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	88641	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	87547	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	39698	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	21366	4.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.80%
7) Chloroethane-d5	1.919	69	20170	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.571	65	12826	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.40%
20) 2-Butanone-d5	4.661	46	110611	55.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.90%
24) Chloroform-d	5.073	84	56032	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.713	65	35611	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
32) Benzene-d6	5.735	84	100386	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.700	67	36197	5.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.00%
41) Toluene-d8	7.906	98	84646	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.80%
43) trans-1,3-Dichloroprop...	8.185	79	12598	4.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.40%
46) 2-Hexanone-d5	8.645	63	66792	44.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.72%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	29121	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	32062	4.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
Data File : VU043423.D
Acq On : 29 Apr 2021 14:27
Operator : SY/MD
Sample : M2197-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YARS9

Quant Time: Apr 30 07:28:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 30 07:27:36 2021
Response via : Initial Calibration

