

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
 Data File : VU043440.D
 Acq On : 29 Apr 2021 21:16
 Operator : SY/MD
 Sample : M2197-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YARH2

Quant Time: Apr 30 07:32:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 30 07:27:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	82025	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	83846	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	37245	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	18774	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.919	69	19909	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.40%
11) 1,1-Dichloroethene-d2	2.572	65	12041	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.80%
20) 2-Butanone-d5	4.639	46	102928	56.26	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.52%
24) Chloroform-d	5.073	84	56768	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
26) 1,2-Dichloroethane-d4	5.710	65	35150	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.735	84	97061	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.700	67	34998	5.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.20%
41) Toluene-d8	7.906	98	78918	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloroprop...	8.189	79	11826	4.14	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.80%
46) 2-Hexanone-d5	8.642	63	63292	44.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.78%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	28203	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	31067	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	4.678	96	4320	0.75	ug/L	90
25) Chloroform	5.096	83	4380	0.39	ug/L	89
31) Carbon tetrachloride	5.542	117	1444	0.17	ug/L	97
34) Trichloroethene	6.549	95	55635	9.01	ug/L	95
47) Tetrachloroethene	8.562	164	13039	2.66	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
Data File : VU043440.D
Acq On : 29 Apr 2021 21:16
Operator : SY/MD
Sample : M2197-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YARH2

Quant Time: Apr 30 07:32:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 30 07:27:36 2021
Response via : Initial Calibration

