

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
 Data File : VU043442.D
 Acq On : 29 Apr 2021 22:29
 Operator : SY/MD
 Sample : VU0430WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK141

Quant Time: Apr 30 08:35:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 30 08:34:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	85306	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	85932	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	37519	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	20330	4.69	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.80%
7) Chloroethane-d5	1.916	69	19508	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.40%
11) 1,1-Dichloroethene-d2	2.571	65	13263	5.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.60%
20) 2-Butanone-d5	4.655	46	112820	59.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.60%
24) Chloroform-d	5.073	84	57797	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.713	65	36229	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	5.735	84	101443	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.700	67	35899	5.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.20%
41) Toluene-d8	7.906	98	82487	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.40%
43) trans-1,3-Dichloroprop...	8.189	79	11963	4.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.60%
46) 2-Hexanone-d5	8.645	63	66112	45.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.48%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	29129	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.80%
66) 1,2-Dichlorobenzene-d4	12.201	152	31670	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%
Target Compounds						
16) Methylene chloride	3.057	84	948	0.15	ug/L	Qvalue 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU043021\
Data File : VU043442.D
Acq On : 29 Apr 2021 22:29
Operator : SY/MD
Sample : VU0430WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK141

Quant Time: Apr 30 08:35:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 30 08:34:29 2021
Response via : Initial Calibration

