

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050118\
 Data File : VU023552.D
 Acq On : 01 May 2018 14:10
 Operator : MD/SY
 Sample : VU0501WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 02 03:24:55 2018
 Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 02 03:12:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	326270	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	317297	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	153518	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	112780	48.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.54%
7) Chloroethane-d5	1.68	69	87486	49.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.26%
11) 1,1-Dichloroethene-d2	2.27	63	159174	35.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.62%
21) 2-Butanone-d5	4.18	46	175910	110.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.10%
24) Chloroform-d	4.65	84	183631	48.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.04%
26) 1,2-Dichloroethane-d4	5.31	65	135252	50.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.70%
32) Benzene-d6	5.35	84	436768	51.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.62%
36) 1,2-Dichloropropane-d6	6.34	67	145016	50.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.72%
41) Toluene-d8	7.57	98	395791	50.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.86%
43) trans-1,3-Dichloropropene-	7.85	79	60805	49.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.06%
47) 2-Hexanone-d5	8.31	63	115176	106.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.36%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	180857	49.03	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.06%
64) 1,2-Dichlorobenzene-d4	11.86	152	156625	52.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050118\
Data File : VU023552.D
Acq On : 01 May 2018 14:10
Operator : MD/SY
Sample : VU0501WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 02 03:24:55 2018
Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 02 03:12:05 2018
Response via : Initial Calibration

