

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050123\
 Data File : VU054026.D
 Acq On : 01 May 2023 15:19
 Operator : JC/MD
 Sample : 02525-08ME
 Misc : 5.68g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCCK5ME

Quant Time: May 02 05:33:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 02 05:29:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	241983	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.411	117	234232	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	113041	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	76457	41.777	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.560%
7) Chloroethane-d5	1.909	69	53336	44.397	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.800%
11) 1,1-Dichloroethene-d2	2.562	63	109591	33.762	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.520%
21) 2-Butanone-d5	4.620	46	118413	113.444	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.440%
24) Chloroform-d	5.057	84	163865	48.751	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.500%
26) 1,2-Dichloroethane-d4	5.694	65	112970	52.346	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.700%
32) Benzene-d6	5.719	84	324093	47.007	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.020%
36) 1,2-Dichloropropane-d6	6.684	67	108465	49.833	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.660%
41) Toluene-d8	7.893	98	287642	46.210	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.420%
43) trans-1,3-Dichloroprop...	8.173	79	52839	49.055	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.100%
47) 2-Hexanone-d5	8.629	63	78709	100.962	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.960%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	144467	50.390	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.780%
66) 1,2-Dichlorobenzene-d4	12.189	152	116602	52.802	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050123\
Data File : VU054026.D
Acq On : 01 May 2023 15:19
Operator : JC/MD
Sample : 02525-08ME
Misc : 5.68g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCCK5ME

Quant Time: May 02 05:33:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 02 05:29:54 2023
Response via : Initial Calibration

