

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU050219\
 Data File : VU031708.D
 Acq On : 02 May 2019 13:35
 Operator : JC/SP
 Sample : K2633-19 50X
 Misc : 5.01g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAJB1

Quant Time: May 03 06:26:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 03 04:37:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	1008730	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	947589	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	444269	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	210392	24.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	49.84%#
7) Chloroethane-d5	1.68	69	191379	29.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	58.66%#
11) 1,1-Dichloroethene-d2	2.27	63	345199	20.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	41.06%#
21) 2-Butanone-d5	4.19	46	729057	95.31	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.31%
24) Chloroform-d	4.64	84	524580	33.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.86%#
26) 1,2-Dichloroethane-d4	5.31	65	348814	34.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	69.06%#
32) Benzene-d6	5.34	84	1021454	35.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.20%
36) 1,2-Dichloropropane-d6	6.32	67	394076	38.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	76.98%
41) Toluene-d8	7.56	98	894724	35.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.60%#
43) trans-1,3-Dichloropropene-	7.85	79	168998	38.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.92%
47) 2-Hexanone-d5	8.31	63	500795	113.02	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.02%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	563332	41.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	372291	39.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.70%#

Target Compounds

					Ovalue
29) Cyclohexane	4.99	56	17243	1.150 ug/L	# 91
33) Benzene	5.39	78	244115	7.007 ug/L	100
35) Methylcyclohexane	6.41	83	22221	1.756 ug/L	98
42) Toluene	7.63	91	461998	13.312 ug/L	99
52) Ethylbenzene	9.25	91	224857	6.246 ug/L	99
53) m,p-Xylene	9.37	106	644334	51.372 ug/L	97
54) o-xylene	9.77	106	294188	23.659 ug/L	94
55) Styrene	9.79	104	859736	40.469 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU050219\
Data File : VU031708.D
Acq On : 02 May 2019 13:35
Operator : JC/SP
Sample : K2633-19 50X
Misc : 5.01µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAJB1

Quant Time: May 03 06:26:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 03 04:37:29 2019
Response via : Initial Calibration

