

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050222\
 Data File : VU048362.D
 Acq On : 02 May 2022 12:00
 Operator : SY/MD
 Sample : N2638-22
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL27

Quant Time: May 03 04:12:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 03 04:11:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	390640	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	398206	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	189212	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	134695	37.496	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	75.000%	
7) Chloroethane-d5	1.896	69	112981	44.807	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.620%	
11) 1,1-Dichloroethene-d2	2.565	63	186154	32.342	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	64.680%	
21) 2-Butanone-d5	4.620	46	275152	89.807	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	89.810%	
24) Chloroform-d	5.063	84	305340	46.862	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.720%	
26) 1,2-Dichloroethane-d4	5.700	65	191406	49.511	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.020%	
32) Benzene-d6	5.729	84	631209	50.260	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.520%	
36) 1,2-Dichloropropane-d6	6.690	67	205609	49.520	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.040%	
41) Toluene-d8	7.899	98	567767	49.875	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.740%	
43) trans-1,3-Dichloroprop...	8.179	79	89922	49.082	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.160%	
47) 2-Hexanone-d5	8.635	63	204160	91.139	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	91.140%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	343143	47.123	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	94.240%	
66) 1,2-Dichlorobenzene-d4	12.195	152	235871	55.416	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050222\
Data File : VU048362.D
Acq On : 02 May 2022 12:00
Operator : SY/MD
Sample : N2638-22
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL27

Quant Time: May 03 04:12:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 03 04:11:15 2022
Response via : Initial Calibration

