




Data File : VU031758.D
Acq On : 03 May 2019 18:23
Operator : JC/SP
Sample : K2695-04 400X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X89

Manual Integrations
APPROVED

MMDadoda
5/6/2019 11:43:46 AM

Quant Time: May 04 06:58:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 04 02:35:16 2019
Response via : Initial Calibration



TIC: VU031758.D

(24) Chloroform-d (S)

4.875min (+0.232) 0.02ug/L

response 272

Ion	Exp%	Act%
84.00	100	100
86.00	62.50	64.34
47.00	45.70	52.94
49.00	27.20	30.88




Data File : VU031758.D
Acq On : 03 May 2019 18:23
Operator : JC/SP
Sample : K2695-04 400X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X89

Manual Integrations
APPROVED

MMDadoda
5/6/2019 11:43:46 AM

Quant Time: May 04 06:58:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 04 02:35:16 2019
Response via : Initial Calibration



TIC: VU031758.D

(24) Chloroform-d (S)

4.643min (+0.000) 44.45ug/L m

response 524356

Ion	Exp%	Act%
84.00	100	100
86.00	62.50	0.03#
47.00	45.70	0.03#
49.00	27.20	0.02#

Data File : VU031758.D
Acq On : 03 May 2019 18:23
Operator : JC/SP
Sample : K2695-04 400X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X89

Manual Integrations
APPROVED

MMDadoda
5/6/2019 11:43:46 AM

Quant Time: May 04 06:59:47 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 04 02:35:16 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	769645	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	740049	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	284392	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	320453	49.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.50%
7) Chloroethane-d5	1.68	69	250131	50.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.50%
11) 1,1-Dichloroethene-d2	2.27	63	459608	35.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.66%
21) 2-Butanone-d5	4.19	46	632032	108.30	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.30%
24) Chloroform-d	4.64	84	524356m	44.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.90%
26) 1,2-Dichloroethane-d4	5.31	65	367022	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.24%
32) Benzene-d6	5.34	84	1092660	48.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.16%
36) 1,2-Dichloropropane-d6	6.32	67	380184	47.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.10%
41) Toluene-d8	7.56	98	887972	44.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.72%
43) trans-1,3-Dichloropropene-	7.85	79	150244	43.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.56%
47) 2-Hexanone-d5	8.31	63	377338	109.04	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	493702	46.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.02%
64) 1,2-Dichlorobenzene-d4	11.86	152	280208	46.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.70%

Target Compounds						Qvalue
34) Trichloroethene	6.19	95	61932	9.390	ug/L	96
46) Tetrachloroethene	8.23	164	3186	0.709	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU031758.D
Acq On : 03 May 2019 18:23
Operator : JC/SP
Sample : K2695-04 400X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X89

Manual Integrations
APPROVED

MMDadoda
5/6/2019 11:43:46 AM

Quant Time: May 04 06:59:47 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 04 02:35:16 2019
Response via : Initial Calibration

