

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050321\
 Data File : VU043483.D
 Acq On : 03 May 2021 12:49
 Operator : SY/MD
 Sample : VU0503WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK143

Quant Time: May 04 07:39:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 04 07:31:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	94136	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	94867	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	45427	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	39493	5.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.40%
7) Chloroethane-d5	1.919	69	28949	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.571	65	19393	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.40%
20) 2-Butanone-d5	4.652	46	128011	54.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.94%
24) Chloroform-d	5.073	84	69579	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.713	65	46028	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.735	84	134732	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.700	67	46242	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	7.906	98	116445	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
43) trans-1,3-Dichloroprop...	8.189	79	18306	5.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.80%
46) 2-Hexanone-d5	8.645	63	79402	52.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.30%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	35396	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	43300	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050321\
Data File : VU043483.D
Acq On : 03 May 2021 12:49
Operator : SY/MD
Sample : VU0503WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK143

Quant Time: May 04 07:39:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 04 07:31:29 2021
Response via : Initial Calibration

