

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
 Data File : VU048384.D
 Acq On : 03 May 2022 11:14
 Operator : SY/MD
 Sample : VU0503WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK112

Quant Time: May 04 01:38:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 04 01:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	426037	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	442442	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	219264	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	209428	53.457	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.920%		
7) Chloroethane-d5	1.887	69	159908	58.148	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery	= 116.300%		
11) 1,1-Dichloroethene-d2	2.562	63	256518	40.864	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 81.720%		
21) 2-Butanone-d5	4.620	46	300895	90.049	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 90.050%		
24) Chloroform-d	5.063	84	333379	46.915	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 93.820%		
26) 1,2-Dichloroethane-d4	5.700	65	212164	50.321	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.640%		
32) Benzene-d6	5.726	84	726284	52.048	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.100%		
36) 1,2-Dichloropropane-d6	6.690	67	229350	49.716	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 99.440%		
41) Toluene-d8	7.896	98	649410	51.343	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 102.680%		
43) trans-1,3-Dichloroprop...	8.179	79	96523	47.417	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 94.840%		
47) 2-Hexanone-d5	8.636	63	232062	93.237	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 93.240%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	376385	46.520	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 93.040%		
66) 1,2-Dichlorobenzene-d4	12.195	152	263184	53.358	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
Data File : VU048384.D
Acq On : 03 May 2022 11:14
Operator : SY/MD
Sample : VU0503WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK112

Quant Time: May 04 01:38:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 04 01:37:47 2022
Response via : Initial Calibration

