

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
 Data File : VU048399.D
 Acq On : 03 May 2022 18:42
 Operator : SY/MD
 Sample : N2668-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL75

Quant Time: May 04 01:41:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 04 01:37:47 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	431658	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	433707	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	194371	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	211764	53.349	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.700%		
7) Chloroethane-d5	1.900	69	153557	55.112	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 110.220%		
11) 1,1-Dichloroethene-d2	2.565	63	251893	39.605	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 79.220%		
21) 2-Butanone-d5	4.620	46	303865	89.754	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 89.750%		
24) Chloroform-d	5.063	84	343030	47.644	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 95.280%		
26) 1,2-Dichloroethane-d4	5.703	65	215021	50.334	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.660%		
32) Benzene-d6	5.726	84	730329	53.392	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.780%		
36) 1,2-Dichloropropane-d6	6.690	67	235574	52.093	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 104.180%		
41) Toluene-d8	7.899	98	637134	51.387	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 102.780%		
43) trans-1,3-Dichloroprop...	8.179	79	96613	48.417	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 96.840%		
47) 2-Hexanone-d5	8.632	63	212342	87.032	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 87.030%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	362791	45.743	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 91.480%		
66) 1,2-Dichlorobenzene-d4	12.195	152	243313	55.647	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 111.300%		
Target Compounds						
46) Tetrachloroethene	8.555	164	24687	9.139	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
Data File : VU048399.D
Acq On : 03 May 2022 18:42
Operator : SY/MD
Sample : N2668-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL75

Quant Time: May 04 01:41:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 04 01:37:47 2022
Response via : Initial Calibration

