

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
 Data File : VU048391.D
 Acq On : 03 May 2022 15:31
 Operator : SY/MD
 Sample : N2668-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL73

Quant Time: May 04 01:40:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 04 01:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	430809	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	427398	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	197618	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	215497	54.396	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 108.800%		
7) Chloroethane-d5	1.900	69	168646	60.647	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 121.300%		
11) 1,1-Dichloroethene-d2	2.565	63	254918	40.160	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 80.320%		
21) 2-Butanone-d5	4.620	46	300411	88.909	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 88.910%		
24) Chloroform-d	5.063	84	339362	47.228	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 94.460%		
26) 1,2-Dichloroethane-d4	5.700	65	214321	50.269	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.540%		
32) Benzene-d6	5.726	84	727543	53.974	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.940%		
36) 1,2-Dichloropropane-d6	6.690	67	230544	51.734	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 103.460%		
41) Toluene-d8	7.899	98	638478	52.256	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 104.520%		
43) trans-1,3-Dichloroprop...	8.179	79	97824	49.748	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 99.500%		
47) 2-Hexanone-d5	8.632	63	214121	89.057	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 89.060%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	350738	44.876	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 89.760%		
66) 1,2-Dichlorobenzene-d4	12.195	152	242222	54.487	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 108.980%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050322\
Data File : VU048391.D
Acq On : 03 May 2022 15:31
Operator : SY/MD
Sample : N2668-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL73

Quant Time: May 04 01:40:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 04 01:37:47 2022
Response via : Initial Calibration

