

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050323\
 Data File : VU054054.D
 Acq On : 03 May 2023 17:25
 Operator : JC/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5104

Quant Time: May 04 06:03:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:10:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	13059	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.345	117	12084	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.741	152	7678	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	3940	0.413	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	82.000%		
4) 1,2-Dichloroethane-d4	5.661	65	4830	0.492	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
7) 1,2-Dichloropropane-d6	6.639	67	5557	0.459	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	92.000%		
8) Toluene-d8	7.830	98	10836	0.482	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.000%		
10) 1,1,2,2-Tetrachloroeth...	10.679	84	5206	0.487	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	98.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.598	62	4298	0.432	ug/L	99
6) Trichloroethene	6.489	95	6598	0.482	ug/L	98
9) 1,2-Dibromoethane	8.866	107	4280	0.524	ug/L	99
12) 1,2,3-Trichloropropane	10.748	75	3614	0.469	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.924	75	654	0.512	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050323\
Data File : VU054054.D
Acq On : 03 May 2023 17:25
Operator : JC/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5104

Quant Time: May 04 06:03:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:10:40 2023
Response via : Initial Calibration

