

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050323\
 Data File : VU054052.D
 Acq On : 03 May 2023 16:09
 Operator : JC/MD
 Sample : VU0503WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK017

Quant Time: May 04 06:02:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:10:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	15132	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.346	117	14918	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	10605	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	4174	0.377	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	76.000%
4) 1,2-Dichloroethane-d4	5.661	65	4982	0.438	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	88.000%
7) 1,2-Dichloropropane-d6	6.639	67	5755	0.385	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	76.000%
8) Toluene-d8	7.830	98	11503	0.414	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	82.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	5758	0.437	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050323\
Data File : VU054052.D
Acq On : 03 May 2023 16:09
Operator : JC/MD
Sample : VU0503WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK017

Quant Time: May 04 06:02:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:10:40 2023
Response via : Initial Calibration

