

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050418\
 Data File : VU023643.D
 Acq On : 04 May 2018 21:02
 Operator : MD/SY
 Sample : J2630-21
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Quant Time: May 05 01:37:01 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR050418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 05 01:17:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	141743	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	145986	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	64342	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	47542	4.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	1.68	69	39241	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.80%
11) 1,1-Dichloroethene-d2	2.27	63	67910	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	4.20	46	157692	53.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.04%
24) Chloroform-d	4.65	84	92636	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.32	65	52196	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.34	84	183496	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.34	67	64188	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.57	98	150782	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	7.86	79	19120	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.32	63	119054	49.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.88%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	50092	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	11.87	152	62413	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050418\
Data File : VU023643.D
Acq On : 04 May 2018 21:02
Operator : MD/SY
Sample : J2630-21
Misc : 25mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Quant Time: May 05 01:37:01 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR050418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 05 01:17:03 2018
Response via : Initial Calibration

