

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU050420\
Data File : VU038003.D
Acq On : 04 May 2020 15:55
Operator : JC/MD
Sample : VU0504WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0504WBL01

Quant Time: May 05 02:35:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U050420DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 05 02:20:02 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.15	96	88609	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.65	95	30189	0.88	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.00%	
68) 1,2-Dichlorobenzene-d4	12.21	152	29644	0.89	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.00%	

Target Compounds	Ovalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU050420\
Data File : VU038003.D
Acq On : 04 May 2020 15:55
Operator : JC/MD
Sample : VU0504WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0504WBL01

Quant Time: May 05 02:35:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U050420DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 05 02:20:02 2020
Response via : Initial Calibration

