

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
 Data File : VU043511.D
 Acq On : 04 May 2021 11:26
 Operator : SY/MD
 Sample : M2195-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB3H3

Quant Time: May 05 04:52:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 05 04:50:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	104949	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	108449	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	50787	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	30239	3.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.60%
7) Chloroethane-d5	1.919	69	23743	3.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.80%
11) 1,1-Dichloroethene-d2	2.572	65	15140	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	4.645	46	113018	43.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.28%
24) Chloroform-d	5.073	84	61983	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.80%
26) 1,2-Dichloroethane-d4	5.713	65	39153	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
32) Benzene-d6	5.735	84	113102	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.20%
36) 1,2-Dichloropropane-d6	6.700	67	40065	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	77.80%
41) Toluene-d8	7.906	98	96459	3.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.60%
43) trans-1,3-Dichloroprop...	8.185	79	15162	3.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.00%
46) 2-Hexanone-d5	8.642	63	72273	41.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.04%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	31350	3.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	35680	3.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.80%#
Target Compounds						
					Qvalue	
13) Acetone	2.655	43	8770	5.02	ug/L	99
25) Chloroform	5.099	83	3055	0.20	ug/L	96
42) Toluene	7.976	91	5132	0.17	ug/L	96
47) Tetrachloroethene	8.562	164	226207	42.09	ug/L	97
53) m,p-Xylene	9.697	106	770	0.07	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
Data File : VU043511.D
Acq On : 04 May 2021 11:26
Operator : SY/MD
Sample : M2195-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB3H3

Quant Time: May 05 04:52:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 05 04:50:58 2021
Response via : Initial Calibration

