

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
 Data File : VU043521.D
 Acq On : 04 May 2021 15:23
 Operator : SY/MD
 Sample : M2195-15 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB3J3

Quant Time: May 05 04:58:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 05 04:50:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	81866	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	84304	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	38031	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	30566	4.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.80%
7) Chloroethane-d5	1.916	69	24214	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.571	65	15627	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.80%
20) 2-Butanone-d5	4.649	46	115376	56.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.90%
24) Chloroform-d	5.073	84	61649	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.713	65	40502	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
32) Benzene-d6	5.735	84	113303	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.700	67	39986	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.906	98	93990	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
43) trans-1,3-Dichloroprop...	8.189	79	15362	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
46) 2-Hexanone-d5	8.645	63	84532	62.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	124.94%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	34736	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.40%
66) 1,2-Dichlorobenzene-d4	12.201	152	35807	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%
Target Compounds						Qvalue
16) Methylene chloride	3.054	84	6645	0.91	ug/L	90
34) Trichloroethene	6.549	95	2563	0.45	ug/L	89
47) Tetrachloroethene	8.562	164	205388	49.16	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
Data File : VU043521.D
Acq On : 04 May 2021 15:23
Operator : SY/MD
Sample : M2195-15 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB3J3

Quant Time: May 05 04:58:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 05 04:50:58 2021
Response via : Initial Calibration

