

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
 Data File : VU043540.D
 Acq On : 04 May 2021 23:25
 Operator : SY/MD
 Sample : M2195-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: May 05 08:20:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 05 08:11:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	79310	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	82908	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	35522	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	30179	4.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.60%
7) Chloroethane-d5	1.916	69	23623	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.572	65	14975	4.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.60%
20) 2-Butanone-d5	4.645	46	105310	53.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.38%
24) Chloroform-d	5.073	84	61557	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.713	65	39244	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
32) Benzene-d6	5.739	84	109786	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.700	67	39636	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.906	98	88877	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
43) trans-1,3-Dichloroprop...	8.189	79	13855	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.645	63	72044	54.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.28%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	33042	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
66) 1,2-Dichlorobenzene-d4	12.201	152	34249	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%
Target Compounds						
16) Methylene chloride	3.054	84	2870	0.40	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
Data File : VU043540.D
Acq On : 04 May 2021 23:25
Operator : SY/MD
Sample : M2195-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: May 05 08:20:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 05 08:11:40 2021
Response via : Initial Calibration

