

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
 Data File : VU043546.D
 Acq On : 05 May 2021 11:22
 Operator : SY/MD
 Sample : M2195-12 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB3J0

Quant Time: May 05 13:05:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 05 08:11:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	79322	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	84511	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	37033	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	27473	4.31	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.20%	
7) Chloroethane-d5	1.919	69	22278	4.77	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.40%	
11) 1,1-Dichloroethene-d2	2.571	65	13840	4.42	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	88.40%	
20) 2-Butanone-d5	4.649	46	105375	53.21	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.42%	
24) Chloroform-d	5.073	84	59304	4.86	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.20%	
26) 1,2-Dichloroethane-d4	5.713	65	37366	5.08	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.60%	
32) Benzene-d6	5.739	84	104354	4.57	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.40%	
36) 1,2-Dichloropropane-d6	6.700	67	38523	4.80	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.00%	
41) Toluene-d8	7.906	98	86154	4.10	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.00%	
43) trans-1,3-Dichloroprop...	8.189	79	13913	4.30	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.00%	
46) 2-Hexanone-d5	8.645	63	72153	53.19	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	106.38%	
56) 1,1,2,2-Tetrachloroeth...	10.764	84	31094	4.84	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.80%	
66) 1,2-Dichlorobenzene-d4	12.201	152	33182	5.02	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.40%	
Target Compounds						
13) Acetone	2.655	43	1961	1.49	ug/L	69
16) Methylene chloride	3.054	84	3228	0.45	ug/L	91
34) Trichloroethene	6.549	95	350	0.06	ug/L	86
42) Toluene	7.976	91	1923	0.08	ug/L	95
47) Tetrachloroethene	8.562	164	167820	40.07	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050421\
Data File : VU043546.D
Acq On : 05 May 2021 11:22
Operator : SY/MD
Sample : M2195-12 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB3J0

Quant Time: May 05 13:05:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 05 08:11:40 2021
Response via : Initial Calibration

