

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
 Data File : VU048407.D
 Acq On : 04 May 2022 10:46
 Operator : SY/MD
 Sample : VU0504WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK114

Quant Time: May 05 02:17:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 05 02:16:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	371467	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	387026	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	177392	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	195071	57.107	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	114.220%	
7) Chloroethane-d5	1.890	69	143013	59.645	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	119.280%	
11) 1,1-Dichloroethene-d2	2.562	63	230658	42.143	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.280%	
21) 2-Butanone-d5	4.620	46	305749	104.944	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	104.940%	
24) Chloroform-d	5.063	84	311953	50.348	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.700%	
26) 1,2-Dichloroethane-d4	5.700	65	198243	53.926	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.860%	
32) Benzene-d6	5.726	84	669414	54.842	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	109.680%	
36) 1,2-Dichloropropane-d6	6.690	67	220367	54.608	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	109.220%	
41) Toluene-d8	7.899	98	585267	52.897	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.800%	
43) trans-1,3-Dichloroprop...	8.179	79	92338	51.856	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	103.720%	
47) 2-Hexanone-d5	8.636	63	223010	102.429	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	102.430%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	362436	51.210	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	102.420%	
66) 1,2-Dichlorobenzene-d4	12.195	152	230615	57.791	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	115.580%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
Data File : VU048407.D
Acq On : 04 May 2022 10:46
Operator : SY/MD
Sample : VU0504WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK114

Quant Time: May 05 02:17:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 05 02:16:45 2022
Response via : Initial Calibration

