

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
 Data File : VU048411.D
 Acq On : 04 May 2022 12:28
 Operator : SY/MD
 Sample : N2668-17
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL68

Quant Time: May 05 02:18:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 05 02:16:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	401225	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	406957	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	179768	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	204762	55.498	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	111.000%
7) Chloroethane-d5	1.893	69	149370	57.675	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	115.360%
11) 1,1-Dichloroethene-d2	2.562	63	236151	39.946	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.900%
21) 2-Butanone-d5	4.620	46	308303	97.972	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.970%
24) Chloroform-d	5.063	84	324291	48.458	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.920%
26) 1,2-Dichloroethane-d4	5.700	65	199036	50.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.260%
32) Benzene-d6	5.726	84	699046	54.465	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.920%
36) 1,2-Dichloropropane-d6	6.690	67	229275	54.033	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.060%
41) Toluene-d8	7.899	98	600563	51.621	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.240%
43) trans-1,3-Dichloroprop...	8.179	79	89301	47.694	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.380%
47) 2-Hexanone-d5	8.635	63	220829	96.460	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.460%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	365862	49.162	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.320%
66) 1,2-Dichlorobenzene-d4	12.195	152	230457	56.988	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.980%
Target Compounds					Qvalue	
22) 2-Butanone	4.713	43	3128	1.047	ug/L	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
Data File : VU048411.D
Acq On : 04 May 2022 12:28
Operator : SY/MD
Sample : N2668-17
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL68

Quant Time: May 05 02:18:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 05 02:16:45 2022
Response via : Initial Calibration

