

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
 Data File : VU048412.D
 Acq On : 04 May 2022 12:52
 Operator : SY/MD
 Sample : N2668-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL76

Quant Time: May 05 02:18:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 05 02:16:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	416104	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	409520	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	178521	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	205948	53.823	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	107.640%	
7) Chloroethane-d5	1.900	69	150786	56.140	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	112.280%	
11) 1,1-Dichloroethene-d2	2.565	63	242965	39.629	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	79.260%	
21) 2-Butanone-d5	4.626	46	308914	94.656	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	94.660%	
24) Chloroform-d	5.063	84	333468	48.047	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	96.100%	
26) 1,2-Dichloroethane-d4	5.703	65	205392	49.877	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	99.760%	
32) Benzene-d6	5.729	84	722252	55.920	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	111.840%	
36) 1,2-Dichloropropane-d6	6.690	67	232609	54.476	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	108.960%	
41) Toluene-d8	7.899	98	619615	52.926	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	105.860%	
43) trans-1,3-Dichloroprop...	8.179	79	88727	47.091	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	94.180%	
47) 2-Hexanone-d5	8.636	63	215128	93.382	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	93.380%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	348254	46.503	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	93.000%	
66) 1,2-Dichlorobenzene-d4	12.195	152	227600	56.675	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	113.340%	
Target Compounds					Qvalue	
46) Tetrachloroethene	8.555	164	240141	94.150	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
Data File : VU048412.D
Acq On : 04 May 2022 12:52
Operator : SY/MD
Sample : N2668-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL76

Quant Time: May 05 02:18:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 05 02:16:45 2022
Response via : Initial Calibration

