

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
 Data File : VU048409.D
 Acq On : 04 May 2022 11:40
 Operator : SY/MD
 Sample : N2668-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL70

Quant Time: May 05 02:18:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 05 02:16:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	404558	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	396469	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	167748	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	206251	55.441	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 110.880%		
7) Chloroethane-d5	1.903	69	159593	61.115	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 122.240%		
11) 1,1-Dichloroethene-d2	2.565	63	244572	41.030	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 82.060%		
21) 2-Butanone-d5	4.626	46	304985	96.119	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 96.120%		
24) Chloroform-d	5.067	84	330565	48.988	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 97.980%		
26) 1,2-Dichloroethane-d4	5.703	65	206119	51.482	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.960%		
32) Benzene-d6	5.729	84	717015	57.342	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 114.680%		
36) 1,2-Dichloropropane-d6	6.691	67	229683	55.561	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 111.120%		
41) Toluene-d8	7.899	98	609804	53.802	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 107.600%		
43) trans-1,3-Dichloroprop...	8.179	79	88811	48.688	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 97.380%		
47) 2-Hexanone-d5	8.636	63	209748	94.044	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 94.040%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	340619	46.981	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 93.960%		
66) 1,2-Dichlorobenzene-d4	12.195	152	218944	58.021	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 116.040%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050422\
Data File : VU048409.D
Acq On : 04 May 2022 11:40
Operator : SY/MD
Sample : N2668-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL70

Quant Time: May 05 02:18:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 05 02:16:45 2022
Response via : Initial Calibration

