

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050518\
 Data File : VU023671.D
 Acq On : 05 May 2018 17:09
 Operator : MD/SY
 Sample : J2636-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Quant Time: May 06 01:26:52 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR050418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun May 06 01:00:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	145313	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	144960	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	66762	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	45356	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.68	69	38817	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.27	63	65064	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	4.20	46	159569	52.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.66%
24) Chloroform-d	4.65	84	90286	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.31	65	51769	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
32) Benzene-d6	5.35	84	183578	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
36) 1,2-Dichloropropane-d6	6.34	67	64362	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.57	98	152575	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.86	79	19662	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.32	63	122969	51.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.86%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	51313	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.80%
64) 1,2-Dichlorobenzene-d4	11.87	152	64019	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050518\
Data File : VU023671.D
Acq On : 05 May 2018 17:09
Operator : MD/SY
Sample : J2636-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Quant Time: May 06 01:26:52 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR050418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun May 06 01:00:29 2018
Response via : Initial Calibration

