

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU050520\
Data File : VU038010.D
Acq On : 05 May 2020 11:26
Operator : JC/MD
Sample : VU0505WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0505WBL01

Quant Time: May 06 01:06:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U050420DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 05 02:20:02 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.15	96	85220	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.65	95	30353	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	
68) 1,2-Dichlorobenzene-d4	12.20	152	29047	0.91	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.00%	

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU050520\
Data File : VU038010.D
Acq On : 05 May 2020 11:26
Operator : JC/MD
Sample : VU0505WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0505WBL01

Quant Time: May 06 01:06:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U050420DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue May 05 02:20:02 2020
Response via : Initial Calibration

