

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
 Data File : VU043555.D
 Acq On : 05 May 2021 17:26
 Operator : SY/MD
 Sample : M2251-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-GW-558-560

Quant Time: May 05 17:41:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	107094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	191414	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	187048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	84474	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	116108	64.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	128.98%	
35) Dibromofluoromethane	5.298	113	78108	54.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.60%	
50) Toluene-d8	7.906	98	268748	52.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.60%	
62) 4-Bromofluorobenzene	10.639	95	92744	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
Target Compounds						
16) Acetone	2.629	43	2916	2.45	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
Data File : VU043555.D
Acq On : 05 May 2021 17:26
Operator : SY/MD
Sample : M2251-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-GW-558-560

Quant Time: May 05 17:41:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

