

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
 Data File : VU043557.D
 Acq On : 05 May 2021 18:12
 Operator : SY/MD
 Sample : M2251-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-GW-598-600

Quant Time: May 05 18:31:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	107202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.257	114	191216	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.424	117	189767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	85613	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	115836	64.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	128.56%	
35) Dibromofluoromethane	5.298	113	79010	54.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.98%	
50) Toluene-d8	7.906	98	273320	53.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.52%	
62) 4-Bromofluorobenzene	10.639	95	93555	45.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.90%	
Target Compounds						
						Qvalue
16) Acetone	2.630	43	16464	15.24	ug/l	99
17) Carbon Disulfide	2.800	76	3718	1.20	ug/l	98
40) Benzene	5.781	78	1250	0.22	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
Data File : VU043557.D
Acq On : 05 May 2021 18:12
Operator : SY/MD
Sample : M2251-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-GW-598-600

Quant Time: May 05 18:31:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

