

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
 Data File : VU043558.D
 Acq On : 05 May 2021 18:35
 Operator : SY/MD
 Sample : M2251-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-TB-20210428

Quant Time: May 05 18:52:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	107783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.260	114	190789	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	187665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	83462	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	115698	63.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	127.70%	
35) Dibromofluoromethane	5.298	113	77417	54.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.00%	
50) Toluene-d8	7.906	98	270855	53.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.78%	
62) 4-Bromofluorobenzene	10.639	95	92095	45.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.68%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
Data File : VU043558.D
Acq On : 05 May 2021 18:35
Operator : SY/MD
Sample : M2251-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-TB-20210428

Quant Time: May 05 18:52:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

