

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
 Data File : VU043559.D
 Acq On : 05 May 2021 18:58
 Operator : SY/MD
 Sample : M2251-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-EB-20210429

Quant Time: May 05 19:15:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	106350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	190644	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	187668	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	82375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	115277	64.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	128.96%	
35) Dibromofluoromethane	5.298	113	77600	54.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.34%	
50) Toluene-d8	7.906	98	267007	52.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.34%	
62) 4-Bromofluorobenzene	10.639	95	90364	44.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.04%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050521\
Data File : VU043559.D
Acq On : 05 May 2021 18:58
Operator : SY/MD
Sample : M2251-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-EB-20210429

Quant Time: May 05 19:15:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

