

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050522\
 Data File : VU048438.D
 Acq On : 05 May 2022 15:34
 Operator : SY/MD
 Sample : N2710-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXL86

Quant Time: May 06 05:39:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 06 05:30:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	345579	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	337536	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	162421	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	142788	44.932	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.860%	
7) Chloroethane-d5	1.896	69	109449	49.066	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery	=	98.140%	
11) 1,1-Dichloroethene-d2	2.562	63	180658	35.480	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	70.960%	
21) 2-Butanone-d5	4.620	46	250672	92.485	ug/L	-0.02
Spiked Amount	100.000	Range 40 - 130	Recovery	=	92.480%	
24) Chloroform-d	5.063	84	259021	44.937	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.880%	
26) 1,2-Dichloroethane-d4	5.700	65	158973	46.483	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.960%	
32) Benzene-d6	5.726	84	571945	53.727	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.460%	
36) 1,2-Dichloropropane-d6	6.690	67	185716	52.769	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	105.540%	
41) Toluene-d8	7.899	98	494503	51.247	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.500%	
43) trans-1,3-Dichloroprop...	8.179	79	75294	48.484	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.960%	
47) 2-Hexanone-d5	8.632	63	179845	94.715	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	94.710%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	286928	46.485	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	92.980%	
66) 1,2-Dichlorobenzene-d4	12.195	152	201346	55.107	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.220%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050522\
Data File : VU048438.D
Acq On : 05 May 2022 15:34
Operator : SY/MD
Sample : N2710-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXL86

Quant Time: May 06 05:39:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 06 05:30:23 2022
Response via : Initial Calibration

