

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050618\
 Data File : VU023683.D
 Acq On : 06 May 2018 11:26
 Operator : MD/SY
 Sample : VU0506MBL01
 Misc : 5.00µ/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 07 00:51:26 2018
 Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 07 00:51:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	314771	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	307845	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	142540	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	106819	47.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.76%
7) Chloroethane-d5	1.68	69	85619	50.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.70%
11) 1,1-Dichloroethene-d2	2.27	63	158366	36.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.82%
21) 2-Butanone-d5	4.18	46	173000	112.24	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.24%
24) Chloroform-d	4.65	84	192113	52.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.14%
26) 1,2-Dichloroethane-d4	5.31	65	130241	50.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.52%
32) Benzene-d6	5.35	84	401997	48.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.34%
36) 1,2-Dichloropropane-d6	6.33	67	136492	49.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.68%
41) Toluene-d8	7.57	98	357687	46.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.94%
43) trans-1,3-Dichloropropene-	7.85	79	55505	46.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.20%
47) 2-Hexanone-d5	8.31	63	114821	109.29	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.29%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	177996	49.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.48%
64) 1,2-Dichlorobenzene-d4	11.87	152	145940	52.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050618\
Data File : VU023683.D
Acq On : 06 May 2018 11:26
Operator : MD/SY
Sample : VU0506MBL01
Misc : 5.00g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 07 00:51:26 2018
Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 07 00:51:10 2018
Response via : Initial Calibration

