

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050622\
 Data File : VU048459.D
 Acq On : 06 May 2022 16:55
 Operator : SY/MD
 Sample : VU0506MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK117

Quant Time: May 06 23:44:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 06 23:43:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	388627	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	367798	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	163225	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	148448	44.353	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.700%
7) Chloroethane-d5	1.900	69	106254	48.644	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.280%
11) 1,1-Dichloroethene-d2	2.565	63	204549	34.340	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.680%
21) 2-Butanone-d5	4.620	46	271924	95.400	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.400%
24) Chloroform-d	5.063	84	261740	45.024	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.040%
26) 1,2-Dichloroethane-d4	5.700	65	178868	47.333	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.660%
32) Benzene-d6	5.726	84	543927	48.666	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.340%
36) 1,2-Dichloropropane-d6	6.690	67	178226	48.900	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.800%
41) Toluene-d8	7.899	98	453123	45.841	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.680%
43) trans-1,3-Dichloroprop...	8.179	79	75726	45.142	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.280%
47) 2-Hexanone-d5	8.632	63	170919	96.890	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.890%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	260375	46.900	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	156280	48.352	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050622\
Data File : VU048459.D
Acq On : 06 May 2022 16:55
Operator : SY/MD
Sample : VU0506MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK117

Quant Time: May 06 23:44:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 06 23:43:33 2022
Response via : Initial Calibration

