

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050622\
 Data File : VU048460.D
 Acq On : 06 May 2022 17:19
 Operator : SY/MD
 Sample : VU0506WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK118

Quant Time: May 06 23:44:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 06 23:43:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	364557	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	352862	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	161797	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	137741	43.871	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	87.740%	
7) Chloroethane-d5	1.893	69	98614	48.127	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.260%	
11) 1,1-Dichloroethene-d2	2.562	63	193067	34.552	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.100%	
21) 2-Butanone-d5	4.623	46	267158	99.916	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	99.920%	
24) Chloroform-d	5.063	84	247322	45.353	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.700%	
26) 1,2-Dichloroethane-d4	5.703	65	172106	48.551	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.100%	
32) Benzene-d6	5.726	84	512899	47.833	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.660%	
36) 1,2-Dichloropropane-d6	6.690	67	170557	48.777	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	97.560%	
41) Toluene-d8	7.899	98	429757	45.317	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.640%	
43) trans-1,3-Dichloroprop...	8.179	79	74144	46.069	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	92.140%	
47) 2-Hexanone-d5	8.636	63	171971	101.613	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	101.610%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	265042	49.762	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.520%	
66) 1,2-Dichlorobenzene-d4	12.195	152	155147	48.425	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.860%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050622\
Data File : VU048460.D
Acq On : 06 May 2022 17:19
Operator : SY/MD
Sample : VU0506WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK118

Quant Time: May 06 23:44:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 06 23:43:33 2022
Response via : Initial Calibration

