

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050624\
 Data File : VU058829.D
 Acq On : 06 May 2024 13:02
 Operator : MD/SY
 Sample : P2392-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAQ5

Quant Time: May 07 01:10:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 07 01:07:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	74029	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	71472	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	32126	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.589	65	15481	2.659	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	53.200%
7) Chloroethane-d5	1.904	69	12890	2.938	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	58.800%#
11) 1,1-Dichloroethene-d2	2.557	65	10041	3.425	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.400%
20) 2-Butanone-d5	4.615	46	39498	38.607	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.220%
24) Chloroform-d	5.049	84	48971	4.114	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
26) 1,2-Dichloroethane-d4	5.689	65	24907	4.317	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
32) Benzene-d6	5.718	84	85245	3.747	ug/L	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050624\
Data File : VU058829.D
Acq On : 06 May 2024 13:02
Operator : MD/SY
Sample : P2392-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAQ5

Quant Time: May 07 01:10:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 07 01:07:14 2024
Response via : Initial Calibration

