

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050624\
 Data File : VU058825.D
 Acq On : 06 May 2024 11:27
 Operator : MD/SY
 Sample : P2392-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAS1

Quant Time: May 07 01:09:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 07 01:07:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	61482	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	60079	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28102	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.589	65	15853	3.278	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.600%
7) Chloroethane-d5	1.901	69	12796	3.511	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.200%
11) 1,1-Dichloroethene-d2	2.554	65	10198	4.188	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.800%
20) 2-Butanone-d5	4.621	46	39000	45.900	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.800%
24) Chloroform-d	5.049	84	48476	4.903	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.692	65	25368	5.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	5.714	84	84355	4.411	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.682	67	23899	4.190	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.800%
41) Toluene-d8	7.891	98	73737	4.308	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	8.174	79	8870	4.271	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.627	63	25676	41.725	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.460%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	15683	4.514	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	25374	5.181	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050624\
Data File : VU058825.D
Acq On : 06 May 2024 11:27
Operator : MD/SY
Sample : P2392-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAS1

Quant Time: May 07 01:09:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 07 01:07:14 2024
Response via : Initial Calibration

