

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050718\
 Data File : VU023719.D
 Acq On : 07 May 2018 17:43
 Operator : MD/SY
 Sample : J2725-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Quant Time: May 08 05:16:25 2018
 Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 08 04:32:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	392664	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	363964	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	183102	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	93473	33.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.18%
7) Chloroethane-d5	1.68	69	82642	38.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.92%
11) 1,1-Dichloroethene-d2	2.27	63	140406	25.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.76%#
21) 2-Butanone-d5	4.18	46	208014	108.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.18%
24) Chloroform-d	4.65	84	199965	43.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.90%
26) 1,2-Dichloroethane-d4	5.31	65	133598	41.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.48%
32) Benzene-d6	5.35	84	422616	43.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.56%
36) 1,2-Dichloropropane-d6	6.34	67	145362	44.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.88%
41) Toluene-d8	7.57	98	379729	42.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.36%
43) trans-1,3-Dichloropropene-	7.85	79	57075	40.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.06%
47) 2-Hexanone-d5	8.31	63	156994	126.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	126.39%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	200656	47.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.86%
64) 1,2-Dichlorobenzene-d4	11.87	152	163124	45.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050718\
Data File : VU023719.D
Acq On : 07 May 2018 17:43
Operator : MD/SY
Sample : J2725-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Quant Time: May 08 05:16:25 2018
Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050118WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 08 04:32:26 2018
Response via : Initial Calibration

