

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043605.D
 Acq On : 07 May 2021 12:29
 Operator : SY/MD
 Sample : M2333-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-TB-20210503

Quant Time: May 08 01:07:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.382	168	120650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	218005	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	210372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	95242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	123268	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
35) Dibromofluoromethane	5.298	113	85264	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
50) Toluene-d8	7.906	98	299066	48.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.02%	
62) 4-Bromofluorobenzene	10.639	95	105504	44.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.98%	
Target Compounds						
95) Naphthalene	14.095	128	5918	2.23	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043605.D
Acq On : 07 May 2021 12:29
Operator : SY/MD
Sample : M2333-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-TB-20210503

Quant Time: May 08 01:07:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

