

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043612.D
 Acq On : 07 May 2021 15:07
 Operator : SY/MD
 Sample : M2304-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OWL-C2-GW

Quant Time: May 08 01:09:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	115936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	209504	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	203426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	93315	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	121982	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
35) Dibromofluoromethane	5.298	113	82394	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	
50) Toluene-d8	7.906	98	291917	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
62) 4-Bromofluorobenzene	10.639	95	104751	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	4622	3.55	ug/l	92
73) Isopropylbenzene	10.491	105	12995	1.94	ug/l	99
78) n-propylbenzene	10.912	91	5340	0.65	ug/l	97
83) tert-Butylbenzene	11.427	119	2591	0.47	ug/l	86
85) sec-Butylbenzene	11.652	105	17553	3.59	ug/l	99
89) n-Butylbenzene	12.214	91	3923	3.11	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043612.D
Acq On : 07 May 2021 15:07
Operator : SY/MD
Sample : M2304-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OWL-C2-GW

Quant Time: May 08 01:09:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

