

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043614.D
 Acq On : 07 May 2021 15:54
 Operator : SY/MD
 Sample : M2304-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OWL-C5-GW

Quant Time: May 08 01:10:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	125754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	224492	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	220111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	103611	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	128422	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
35) Dibromofluoromethane	5.298	113	85590	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
50) Toluene-d8	7.906	98	308761	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	10.639	95	114106	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	
Target Compounds						
						Qvalue
3) Chloromethane	1.523	50	1796	1.12	ug/l	95
16) Acetone	2.629	43	13564	9.60	ug/l	98
25) 2-Butanone	4.716	43	3400	1.56	ug/l	95
73) Isopropylbenzene	10.491	105	11748	1.58	ug/l	99
78) n-propylbenzene	10.912	91	13556	1.49	ug/l	98
85) sec-Butylbenzene	11.651	105	14346	2.99	ug/l	94
89) n-Butylbenzene	12.217	91	5801	3.30	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043614.D
Acq On : 07 May 2021 15:54
Operator : SY/MD
Sample : M2304-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OWL-C5-GW

Quant Time: May 08 01:10:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

