

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043616.D
 Acq On : 07 May 2021 16:40
 Operator : SY/MD
 Sample : M2304-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OWL-C7-GW

Quant Time: May 08 01:10:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	121068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	214307	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	210194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	98671	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	122451	49.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.92%	
35) Dibromofluoromethane	5.298	113	83146	48.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.30%	
50) Toluene-d8	7.906	98	297883	49.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.32%	
62) 4-Bromofluorobenzene	10.639	95	108417	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
Target Compounds						
16) Acetone	2.633	43	2371	1.74	ug/l	98
73) Isopropylbenzene	10.491	105	6772	0.96	ug/l	97
83) tert-Butylbenzene	11.427	119	3607	0.61	ug/l	86
85) sec-Butylbenzene	11.648	105	7534	2.25	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043616.D
Acq On : 07 May 2021 16:40
Operator : SY/MD
Sample : M2304-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OWL-C7-GW

Quant Time: May 08 01:10:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

