

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043618.D
 Acq On : 07 May 2021 17:26
 Operator : SY/MD
 Sample : M2303-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-18

Quant Time: May 08 01:11:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	126215	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	223294	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	220031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	101238	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	124751	48.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.66%	
35) Dibromofluoromethane	5.298	113	84815	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
50) Toluene-d8	7.906	98	309217	48.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.94%	
62) 4-Bromofluorobenzene	10.639	95	110603	46.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.08%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	3.189	59	2633	3.84	ug/l	# 84
16) Acetone	2.630	43	6700	4.72	ug/l	100
73) Isopropylbenzene	10.488	105	1792	0.25	ug/l	96
78) n-propylbenzene	10.912	91	2310	0.26	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043618.D
Acq On : 07 May 2021 17:26
Operator : SY/MD
Sample : M2303-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-18

Quant Time: May 08 01:11:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

