

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
 Data File : VU043621.D
 Acq On : 07 May 2021 18:36
 Operator : SY/MD
 Sample : M2303-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DRW-2

Quant Time: May 08 01:12:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	130518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	229297	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	218714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	103595	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	126714	47.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.92%	
35) Dibromofluoromethane	5.298	113	87400	47.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.58%	
50) Toluene-d8	7.906	98	317473	48.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.92%	
62) 4-Bromofluorobenzene	10.639	95	116099	47.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.12%	
Target Compounds						
16) Acetone	2.629	43	1928	1.31	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050721\
Data File : VU043621.D
Acq On : 07 May 2021 18:36
Operator : SY/MD
Sample : M2303-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DRW-2

Quant Time: May 08 01:12:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

